



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/001,626	10/24/2001	Kenneth Alley	ALLE-P3.2-US	4440
21616	7590	01/03/2006	EXAMINER	
LAW OFFICES OF MARK A. GARZIA, P.C. 2058 CHICHESTER AVE BOOTHWYN, PA 19061			SIEFKE, SAMUEL P	
			ART UNIT	PAPER NUMBER
			1743	

DATE MAILED: 01/03/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

<b>Office Action Summary</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/001,626	ALLEY, KENNETH	
	<b>Examiner</b>	<b>Art Unit</b>	
	Samuel P. Siefke	1743	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

#### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) ☒ Responsive to communication(s) filed on 10/14/05.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

#### Disposition of Claims

- 4) ☒ Claim(s) 1-4 and 7-10 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-4 and 7-10 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

#### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)   | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                   | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)             |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 112***

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 1-4,7-10 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. A plurality of overlapping layers is not disclosed in the specification. Webster's Ninth New Collegiate Dictionary defines overlapping as "to extend over and cover a part of." The only mention of a overlapping is on the bridge of page 15 to 16, "Although the test strip material (preferably without the chemically reactive portion) may extend vertically to the bottom of the inner cavity 39, it is preferable to have the test strip overlap an absorbent wick 78." The Examiner argues that the wick 78 is not part of the test strip.

Claims 1-4,7-10 are rejected under 35 U.S.C. 112, first paragraph, as based on a disclosure which is not enabling. The collar 52 is critical or essential to the practice of the invention, but not included in the claim(s) is not enabled by the disclosure. See *In re Mayhew*, 527 F.2d 1229, 188 USPQ 356 (CCPA 1976). This feature is critical because

Art Unit: 1743

the collar is the reason the test strip does not separate. The collar is designed to apply controlled pressure on the test strip...The pressure is required to improve the functionality of the test strip" This excerpt was taken from page 15 of Applicant's specification. Further the test strip does not contain any features that prevents the separation of the layers.

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-4 and 8-10 are rejected under 35 U.S.C. 102(e) as being anticipated by Knappe et al. (USPN 6,537,496).

Knappe discloses a flat-shaped functional overlay for use with a test strip. This overlay provides the test strip to bend (creating multiple planes, fig. 5) and adjust to the surface for assaying in order to give a flush contact (pressure points, entire surface) with the surface (col. 7, lines 29-37; fig. 2-7). Knappe inherently has elongated pores when the test strip is bent or flexed because when a test strip is bent or flexed it is

Art Unit: 1743

stretched which creates elongated pores (col. 4, lines 64-65). As seen in figures 5-7 there is no separation of the layers in the test strip.

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claim 7 is rejected under 35 U.S.C. 103(a) as being unpatentable over Knappe et al. (USPN 6,537,496) in view of Porter et al. (USPN 5,709,838).

Knappe discloses a flat-shaped functional overlay for use with a test strip as seen above.

Knappe does not teach a test strip that has a tapered end.

Porter discloses a single use sampling and sample delivery method that comprise a tapering the end of a test strip in order to provide simple sample uptake (fig.

1). It would have been obvious to one of ordinary skill in the art to modify Knappe to include a tapered end in order to provide easy sample delivery to the detection layers.

### ***Response to Arguments***

Applicant's arguments with respect to claims 1-4 and 7-10 have been considered but are moot in view of the new ground(s) of rejection.

Applicant states, "Applicant submits that one skilled in the art after reading both the Knappe patent and Applicant's description would readily understand that these documents describe the identical problem." Knappe discloses the remedy to this deficiency (separation of layers) and that is why the art is used in rejection the instant application.

Applicant states, "Applicant is at a loss as to why the Examiner, who has experience in the field of test strips, is having trouble understanding the concept of the separation of layers - especially when this is a well-known result when test strips are bent." The Examiner is at a loss as to why the Applicant, who has experience in the field of test strips, is having trouble understanding why the test strip of Knappe is any different than in the instant application when bent.

### ***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP

Art Unit: 1743

§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Samuel P. Siefke whose telephone number is 571-272-1262. The examiner can normally be reached on M-F 7:00am-5:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on 571-272-1700. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.


Art Unit: 1743

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Sam P. Siefke



December 22, 2005



Jill Warden  
Supervisory Patent Examiner  
Technology Center 1700